LABEL IN PART: (Btl.) "100 Water Purification Tablets FOR PURIFYING DRINKING WATER IN CANTEENS HALAZONE N. N. R. * * * Each tablet contains 0.004 Gm. (1/16 grain) of Halazone with sodium carbonate, sodium chloride and boric acid."

LIBELED: 11-30-56, N. Dist. Calif.

CHARGE: 501 (b)—the article purported to be and was represented as "Halazone Tablets," a drug the name of which is recognized in the National Formulary, an official compendium; and its strength, while held for sale, differed from the standard set forth in such compendium.

DISPOSITION: 12-11-56. Default—destruction.

5232. Halazone tablets. (F. D. C. No. 39672. S. No. 37-306 M.)

QUANTITY: 188 cases, 300 btls. each, at Brooklyn, N. Y.

SHIPPED: 8-23-55, from Topeka, Kans.

LABEL IN PART: (Btl.) "Tablets Water Purification Individual (Halazone) 100 Tablets * * * Each tablet contains: P-Sulfonedichloramido Benzoic Acid (1/16 Grain) Sodium Carbonate Sodium Chloride Boric Acid" or "100 Water Purification Tablets For Purifying Drinking Water in Canteens * * * Halazone N. N. R. (P-sulfonedichloramido-benzoic acid) Each tablet contains 0.004 Gm. (1/16 grain) of Halazone with sodium borate and chloride."

RESULTS OF INVESTIGATION: Analysis showed that the tablets contained from 44 percent to 91 percent of the declared amount of halazone. The National Formulary provides that *halazone tablets* contain not less than 90 percent of the labeled amount of halazone.

LIBELED: 11-9-56, E. Dist. N. Y.

CHARGE: 501 (b)—the article purported to be and was represented as "Halazone Tablets," a drug the name of which is recognized in the National Formulary, an official compendium, and its strength, while held for sale, differed from the standard set forth in such compendium.

DISPOSITION: 1-11-57. Default—destruction.

5233. Plastic adhesive strips. (F. D. C. No. 39671. S. No. 44-918 M.)

QUANTITY: 60 12-pkg. ctns. at Washington, D. C.

SHIPPED: 9-4-56, from New Rochelle, N. Y., by C. I. Lee & Co., Inc.

LABEL IN PART: (Pkg.) "10¾ x 3 New Plastic Adhesive Strips Washable Waterproof Individually Wrapped Sterilized."

RESULTS OF INVESTIGATION: Examination showed that the article was not sterile but was contaminated with viable micro-organisms.

LIBELED: 11-5-56, Dist. Columbia.

CHARGE: 501 (b)—the article purported to be and was represented as a drug, "Adhesive Absorbent Bandage," the name of which is recognized in the United States Pharmacopeia, an official compendium; and, when shipped, its quality and purity fell below the standard set forth in such compendium.

DISPOSITION: 12-5-56. Default—destruction.

5234. Clinical thermometers. (F. D. C. No. 39450. S. Nos. 48-862/3 M.)

QUANTITY: 28 doz. *oral thermometers* and 19 doz. *rectal thermometers* at Milwaukee, Wis.